Specification: V3



SHENZHEN COMEN MEDICAL INSTRUMENTS CO.,LTD

No.2 of FIYATA Timepiece Building, Nanhuan Avenue, Gongming sub-district, Guangming New District, Shenzhen, P.R. China

Tel: +86-755-26408879 Fax: +86-755-26431232 Email: info@szcomen.com Web: www.comen.com



Ventilator

V3 (Vision.B00)



Technical Specification

•	
Physical Characteristics	
Dimensions	1389mm × 528 mm × 697 mm
	(Height \times Width \times Depth) (including
	trolley);
	343.5mm × 312.5 mm × 258 mm
	(Height \times Width \times Depth)
	(excluding trolley)
Weight	60 kg (with all safe working load)
	10 kg (main unit)
Air Supply	Integrated ultra-silent turbine
Screen	
Screen Size:	12.1" TFT touch screen
Resolution	1280 × 800
Brightness:	Adjustable
Ventilator Specification	
Ventilation mode	V-A/C (Volume assist/control)
	P-A/C (Pressure assist/control)
	V-SIMV (Volume - Synchronized
	Intermittent Mandatory
	Ventilation)
	P-SIMV (Pressure - Synchronized
	Intermittent Mandatory
	Ventilation)
	CPAP/PSV,
	DuoVent,
	APRV,
	PRVC
	PRVC-SIMV
	VS
	PSV-S/T
	HFNC
Invasive Mode	V-A/C, P-A/C, V-SIMV, P-SIMV,
	CPAP/PSV, DuoVent, PRVC, APRV,
	PRVC-SIMV, VS
Non-invasive Mode	P-A/C, P-SIMV, CPAP/PSV,

DuoVent, APRV, F	PSV-S/	T'
------------------	--------	----

Controlled parameter ra	anges
O ₂ %:	21 - 100% (increments of 1 %)
TV (Tidal Volume):	Adult: 100 - 2200 mL (increments of 10 mL) Pediatric: 20 - 300 mL (increments
	of 1 mL)
Respiratory Rate (RR):	1 - 100 bpm (increments of 1 bpm)
fSIMV (Ventilation frequency in SIMV mode):	1 - 60 bpm (increments of 1 bpm).
I:E range:	4:1~1:10.
Tinsp (Inspiratory time):	0.10 - 10 s (increments of 0.05 s).
Tslope (Time of Pressure Rising):	0 - 2.00 s (increments of 0.05 s).
High Pressure Time (Thigh):	0.2 - 30 s (increments of 0.1 s)
Tlow (Low Pressure Time):	0.2 - 30 s (increments of 0.1 s)
Max inspiratory Time (Timax):	0.20 ~ 15.00 s(increments of 0.1 s)
Tpause:	5 % - 60 % (increments of 5 %), Off
ΔPinsp(Inspiratory pressure):	5 - 80 cmH ₂ O (increments of 1 cmH ₂ O)
ΔPsupp:	$0 - 80 \text{ cmH}_2\text{O}$ (increments of 1 cmH ₂ O)
Phigh (High Pressure Level):	0 - 80 cmH ₂ O (increments of 1 cmH ₂ O)
Plow (Low Pressure	0 - 50 cmH ₂ O (increments of 1
Level):	cmH₂O)
PEEP:	0 - 50 cmH ₂ O (increments of 1 cmH ₂ O), Off
Flow trigger	0.5 -15 L/min (increments of 0.1 L/min), Off
Pressure trigger	-10 to - 0.5 cmH₂O (increments of



	0.5 cmH ₂ O), Off	RCexp (Expiratory Time	0 - 10 s
Sign pressure	0-45	Constant)	
Exp % (Expiration termination/trigger	10 - 85% (increments of 5%), Auto	Waveforms	Airway pressure - time, Flow - time, Volume - time
level)		Weaning indicator	
Apnea Ventilation		P0.1	-20 - 0 cmH₂O
TVapnea	Adult: 100 - 2200 mL (increments of 10 mL) Pediatric: 20 - 300 mL	NIF (Maximum negative inspiratory pressure)	-45 - 0 cmH₂O
ΔPapnea	(increments of 1 mL) 5 - 80 cmH₂O (increments of 1	RSBI (Rapid Shallow Breathing Index)	0 - 999 /(L•min)
_ 0,000	cmH ₂ O)	Special Function	
RRapnea (Apnea	1 - 80 bpm (increments of 1 bpm)	Manual Breath	
Respiratory Rate)		Expiration Hold	
Apnea Tinsp	0.20 - 10 s (increments of 0.05 s)	Inspiration Hold	
Sigh		Nebulizer	
Sigh Switch	ON, Off	O2个(O2 enrichment)	
Interval	20 s - 180 min (increments of 1 s	Sputum Suction	
	from 20 to 59 s, increments of 1	Pulmonary View	
	min from 1 to 180 min)	Lung Recruitment Tool	Sustained Insufflation
Cycles Sigh	1 - 20 (increments of 1)	PEEPi monitoring	
Δint.PEEP	1 - 45 cmH₂O (increments of 1 cmH₂O), Off	P-V Tool	Paw - Volume, Flow - Volume, Paw - Flow
Automatic Tube Resista	nce Compensation	Tube Resistance	TDC
Tube Type	endotracheal intubation and	Compensation	TRC
	tracheotomy tube	Smart Sync	IntelliSynTec
Tube I.D.	Adult: 5.0 - 12.0 mm (increments of	O2 Therapy	2-60 L/min
	0.5 mm) Pediatric: 2.5 - 8.0 mm		EtCO2, Vdaw, VDaw/Tve, Vtalv,
	(increments of 0.5 mm)	CO2	V'alv, SlopeCO2, V'CO2, VeCO2,
Compensate	1 -100 % (increments of 1 %) off		ViCO2
Expiration Compensation	ON, Off	Control accuracy	
Switch		02 %	± (3 vol.% +1 % of setting)
Monitoring		TV	± (10 mL + 10% of the set value)
Airway pressure range	Ppeak, Pplat, Pmean (Range -20 - 120 cmH₂O)	Tinsp	± 0.1 s or ± 10 % of setting, whichever is greater
PEEP	$0{\sim}$ 120 cmH2O	I: E	1:4~2:1: ± 10% of the set value;
Tidal volume range:	0∼4000 mL		Other range: ± 15% of the set
Respiratory Rate	ftotal, fmand, fspn (Range 0 - 200 bpm)	RR	value. ±1 bpm
Minute volume range	MV, MVspn, MVleak (Range 0 –	fSIMV	±1 bpm
. 0-	100.0 L/min)	Tslope (Rising Time)	± (0.2s + 20% of the set value)
Resistance	Rinsp, Rexp (0 - 600 cmH ₂ O/L/s)	Phigh	± (2 cmH2O + 5% of the set value)
Compliance	Cstat, Cdyn (0 - 300 mL/cmH ₂ O)	Plow	± (2 cmH2O + 5% of the set value)
Inspired Oxygen (FiO ₂)	15 - 100 %	Thigh	±0.2s or ± 10% of the set value,
WOB (Work of Breathing)	0 – 100.0 J/min		whichever is larger



Tlow	±0.2s or ± 10% of the set value, whichever is larger		cmH2O \sim 300 mL/ cmH2O, \pm (2 mL/ cmH2O + 20% of the actual
Pressure Trigger	± (1 cmH2O + ± 10% of the set		reading).
	value)	RSBI	Within the range of 0 $$ /(min·L) \sim
Flow Trigger	± (1 L/min + 10% of the set value)		999 /(min·L) , ± (3 /(min·L)+15%
Δint.PEEP	2-45cmH2O± (2 cmH2O + 5% of the		of the actual reading).
	set value) (exclude 2)1-2cmH2O±	WOB	Within the range of 0.0 J/min \sim
	(1% of the set value)		100.0 J/min, ± (1 J/min + 15% of
Exp %	±10% (absolute error)		the actual reading).
Fapnea (Apnea	±1bpm	NIF	Within the range of -45.0 cmH2O~
Frequency)			0.0 cmH2O, ± (2 cmH2O + 4% of
ΔPapnea	± (2 cmH2O + 5% of the set value)	DO 4	the actual reading)
TVapnea	± (10 mL + 10% of the set value)	P0.1	Within the range of -20.0 cmH20 \sim
Apnea Tinsp	±0.1s or ± 10% of the set value, whichever is larger		0.0 cmH2O, \pm (2 cmH2O + 4% of the actual reading).
Monitoring Accuracy		RCexp	Within the range of $0.0s\sim10.0s$,
Airway pressure (Ppeak,	Within the range of -20cmH2O \sim		\pm (0.2s + 20 % of the actual
Pplat, Pmean, PEEP)	120 cmH2O, ± (2 cmH2O + 4% of		reading).
	the actual reading)	Alarm Settings	
Tidal Volume (TVi, TVe,	Within the range of 0 mL \sim 100 mL,	Tidal Volume	Upper alarm limit Adult:
TVe/IBW, TVe spn)	± (10 mL + 3% of the actual		110~4000 mL, OFF
	reading);		Pediatric: 25~600 mL, OFF
	Within the range of 100 mL∼4000		Lower alarm limit Adult: OFF, 50~3995 mL
	mL, ± (3 mL + 10% of the actual		Pediatric: OFF, 10~595 mL
Minute Volume (MV,	reading) Within the range of 0.0 L/min \sim	Minute Volume	Upper alarm limit: Adult: 0.2~100.0
MVspn, MVleak)	100.0 L/min, ± (0.2 L/min + 10% of		L/min
wwspii, www.cakj	the actual reading)		Pediatric: 0.2~60.0 L/min
Frequency (ftotal,	Within the range of 0 bpm \sim 200		Lower alarm limit: Adult: 0.1~50.0
fmand, fspn)	bpm, ±1 bpm or ±5% of the actual		L/min
mana, ispin,	reading, whichever is larger		Pediatric: 0.1~30.0 L/min
Inspired Oxygen (FiO ₂)	Within the range of 15 vol.% \sim 100	FiO2(LPO)	Upper alarm limit: 20 vol.% \sim 100
- 1- 1- 1- 1- 1- 1- 1- 1- 1- 1- 1- 1- 1-	vol.%, ± (2.5 vol. % + 2.5% of the		vol.%
	actual reading).		Lower alarm limit:18 vol.% $\sim~98$
Resistance	Within the range of 0		vol.%
	cmH2O/(L/s) \sim 5 cmH2O/(L/s), the	FiO2(HPO)	Upper alarm limit: Min (Oxygen
	accuracy is not defined;		concentration setting value + max
	Within the range of 5		(7 vol.%, oxygen concentration
	cmH2O/(L/s) \sim 20 cmH2O/(L/s),		setting value x
	±10 cmH2O/(L/s);		10%), 100 vol.%) (rounded)
	Within the range of 20		Lower alarm limit:Max (18 vol.%,
	cmH2O/(L/s) \sim 500 cmH2O/(L/s)		oxygen concentration setting
	$(\mbox{exclude}20)$, $\pm50\%$ of the actual		value-max (7 vol.%, oxygen
	reading).		concentration
Compliance	Within the range of 0 mL/		setting value x 10%)) (rounded)



			CO2 accuracy (under all
Paw	Upper alarm limit: 10~90 cmH2O.		conditions):
	Lower alarm limit: OFF,5~ (upper		a) Within the range of 0
	alarm limit -5) cmH2O		mmHg \sim 114 mmHg, \pm (2.25
ftotal	Upper alarm limit: 2~160 bpm,		mmHg $+$ 4% of the reading).
	OFF.		b) Within the range of 115
	Lower alarm limit: OFF,1~159 bpm		mmHg \sim 190 mmHg, the accuracy
Apnea alarm time	5-60 S		is not defined
SideStream CO2		Waveforms	EtCO ₂ - time
module		Resolution	
Displayed numerics	EtCO ₂	Sampling rate and	Comen SideStream: sampling rate:
Measurement RangeCome	en SideStream: 0 mmHg \sim 150 mmHg,	accuracy	50 mL/min; sampling rate control
0% \sim 19.7%, 0 kPa \sim 20 kP	Pa (at 760 mmHg)		accuracy: ± 10mL/min;
Respironics Capno SideStr	ream: 0 mmHg \sim 99 mmHg, 0.0 % \sim		Respironics Capno SideStream:
13.0 %, 0 kPa∼13.2 kPa (a	at 760 mmHg)		sampling rate: 50 mL/min;
Masimo ISA Capno SideSt	ream: 0 mmHg \sim 190 mmHg, 0		sampling rate control accuracy: ±
vol% \sim 25 vol% (at 760	mmHg)		10 mL/min.
Measurement accuracy	Comen SideStream:		Masimo ISA Capno SideStream:
	a) Within the range of 0 mmHg \sim		sampling rate: 50mL/min; sampling
	40 mmHg, ± 2 mmHg;		rate control accuracy: ± 10 mL/min.
	b) Within the range of 41 mmHg \sim	System response time	Masimo mainstream: < 1 s; Masimo
	70 mmHg, ± 5% of the reading;		ISA Capno sidestream: < 3s (use a 2
	c) Within the range of 71 mmHg \sim		m sampling line)
	100 mmHg, ± 8% of the reading;	Rise time (Response	Masimo mainstream: < 1 s; Masimo
	d) Within the range of 101	time)	ISA Capno sidestream: < 3s (use a 2
	mmHg \sim 150 mmHg, ± 10% of the		m sampling line)
	reading.	EtCO₂ Alarm Upper	Comen sidestream: (lower alarm
	Respironics Capno SideStream:	Limits	limit + 2 mmHg) \sim 150 mmHg
	(Note: the gas temperature is 25 $^{\circ}\mathrm{C}$,		Respironics Capno sidestream:
	if respiratory rate is greater than 80		(lower alarm limit + 2 mmHg) \sim 99
	rpm, the accuracy is 12% of the		mmHg
	reading):		Masimo ISA Capno sidestream:
	0 mmHg \sim 38 mmHg: \pm 2 mmHg,		(lower alarm limit + 2 mmHg) \sim
	39 mmHg \sim 99 mmHg: \pm 10% of the		190 mmHg
	actual reading.	EtCO₂ Alarm Lower	Comen sidestream:0 mmHg \sim
	Masimo ISA Capno SideStream:	Limits	(upper alarm limit - 2 mmHg)
	CO2 accuracy (under the condition:		Respironics Capno sidestream: 0
	22°C ±5°C 1013 ± 40 hPa; gas		mmHg \sim (upper alarm limit $$ - $$ 2
	mixture of CO2 and N2.)		mmHg)
	a) Within the range of 0		Masimo ISA Capno sidestream: 0
	mmHg \sim 114 mmHg, \pm (1.52		mmHg \sim (upper alarm limit $$ - $$ 2
	mmHg + 2% of the reading).		mmHg)
	b) Within the range of 114	MainStream CO₂	
	mmHg \sim 190 mmHg, the accuracy	Module	
	is not defined.	Displayed numerics	EtCO ₂



(during non-motion state); Within

	geComen mainstream: 0 mmHg \sim 150 \sim 20 kPa (at 760 mmHg);		mmHg \sim 190 mmHg, the accuracy is not defined;	
Respironics CAPNOSTAT	5: 0 mmHg \sim 150 mmHg, 0% \sim 19.7%,	Resolution		
0 kPa~20 kPa (at 760 m	mHg);	Waveforms	EtCO ₂ - time, V - CO ₂	
Masimo IRMATM mainst	ream: 0 mmHg \sim 190 mmHg, 0	EtCO₂ Alarm Upper	Comen mainstream: (lower alarm	
vol% \sim 25 vol% (at 760	mmHg);	Limits	limit + 2 mmHg) \sim 150 mmHg	
EtCO₂ Measurement	Comen mainstream:		Respironics CAPNOSTAT 5	
Accuracy	a) Within the range of 0mmHg \sim 40mmHg, \pm 2 mmHg;		mainstream: (lower alarm limit	
			+2mmHg) \sim 150 mmHg	
	b) Within the range of 41mmHg \sim		Masimo IRMATM mainstream:	
	70mmHg, \pm 5% of the reading;		(lower alarm limit + 2 mmHg) \sim	
	c) Within the range of 71mmHg \sim		190 mmHg	
	100mmHg, \pm 8% of the reading;	EtCO₂ Alarm Lower	Comen mainstream: 0 mmHg \sim	
	d) Within the range of 101mmHg \sim	Limits	(upper alarm limit - 2 mmHg)	
	150mmHg, $~\pm~$ 10% of the reading.		Respironics CAPNOSTAT 5	
	Respironics CAPNOSTAT 5		mainstream: 0 mmHg \sim (upper	
	mainstream: CO2 accuracy (Note:		alarm limit - 2 mmHg)	
	Temperature :35 $^{\circ}$ C):		Masimo IRMATM mainstream: 0	
	a) Within the range of 0		mmHg \sim (upper alarm limit - 2	
	mmHg \sim 40 mmHg, \pm 2 mmHg;		mmHg)	
	b) Within the range of 41	SpO₂ module:		
	mmHg \sim 70 mmHg, $~\pm~$ 5% of the	Display	Pulse rate (PR)	
	reading;		waveform/parameter, SpO2	
	c) Within the range of 71	SpO2 measurement rangeNellcor SpO2: 0% \sim 100%		
	mmHg \sim 100 mmHg, \pm 8% of the	Masimo SpO2: 1% \sim 100%		
	reading; d) Within the range of 101 mmHg \sim 150 mmHg, \pm 10% of	Comen SpO2: 0% \sim 100%	%	
		SpO2 accuracy	Nellcor SpO2: Within the range of	
			70% \sim 100%, Adult/Pediatric	
	the reading.		measurement accuracy is $~\pm 2\%$	
	Masimo IRMATM mainstream: CO2		(during non-motion state); Within	
	accuracy (under the condition:		the range of 0% \sim 69%,	
	22°C ± 5°C 1013 ± 40 hPa;		measurement accuracy is not	
	gas mixture of CO2 and N2.):		defined.	
	a) Within the range of 0 mmHg \sim		Masimo SpO2: Within the range of	
	114 mmHg, ± (1.52 mmHg + 2%		70% \sim 100%, Adult/Pediatric	
	of the reading);		measurement accuracy is $\pm 2\%$	
	b) Within the range of 114		(during non-motion state), $\pm 3\%$	
	mmHg \sim 190 mmHg, the accuracy		(during motion state); Within the	
	is not defined;		range of 1% \sim 69%, the	
	CO2 accuracy (under all		measurement accuracy is not	
	conditions):		defined.	
	a) Within the range of 0 mmHg \sim		Comen SpO2: Within the range of	
	114 mmHg, \pm (2.25 mmHg + 4%		70%~100%, Adult/ Pediatric	
	of the reading);		measurement accuracy is $\pm 2\%$	

b) Within the range of 114



	the range of 0% \sim 69%, the		bpm) \sim 300 bpm	
	measurement accuracy is not		Masimo SpO2: (Lower alarm limit	
	defined.		+1 bpm) \sim 240 bpm	
PR measurement range	Nellcor SpO2: 20 bpm \sim 300 bpm		Comen SpO2: (Lower alarm limit +1	
	Masimo SpO2: 25 bpm \sim 240 bpm		bpm) \sim 254 bpm	
	Comen SpO2: 20 bpm \sim 300 bpm	Lower PR alarm limit	Nellcor SpO2: 25bpm \sim (Upper	
PR measurement	Nellcor SpO2: resolution: 1 bpm		alarm limit -1bpm)	
resolution	Masimo SpO2: resolution: 1 bpm		Masimo SpO2: 25bpm \sim (Upper	
	Comen SpO2: resolution: 1 bpm		alarm limit -1bpm)	
PR measurement	Nellcor SpO2: 20 bpm \sim 250 bpm:		Comen SpO2: 20bpm \sim (Upper	
accuracy	the measurement error should be		alarm limit -1bpm)	
	\pm 3 bpm; 251 \sim 300 bpm:	Trend		
	measurement accuracy is not	Туре	Tabular, Graphic	
	defined.	Length72 hours		
	Masimo SpO2: the measurement	Content	Monitor Parameters, Setting	
	error should be ±3 bpm (during	-	Parameters (Setting Ventilation	
	non-motion state) and ±5 bpm		mode and Parameters) includes	
	(during motion state)		parameter alarm events and	
	Comen SpO2: the measurement		parameter waveforms related to	
	error should be ±2 bpm		the alarm time	
Perfusion index range	Nellcor SpO2: / (Note: Nellcor SpO2	Data Review	the diarm time	
	module has no perfusion index.)		Lin to 2000 avent land on he	
	Masimo SpO2: $0.02\%\sim20\%$, the	Event logs	Up to 8000 event logs can be	
	accuracy is not defined.		saved, including alarm logs and	
	Comen SpO2: $0.05\%\sim$ 20%, the		operation logs. The alarm log	
	accuracy is not defined.		includes parameter alarm events	
Data and data manifed	·		and parameter waveforms related	
Data update period	≤ 2 s		to the alarm time.	
Signal Quality Index (SIQ)	Masimo SpO2 and Comen SpO2	Freeze the waveform rev	Freeze the waveform reviewFreeze the waveform of the	
ndication function	should come with SIQ indication	interface at the current time and use the knob to review the		
		interface at the current t	ime and use the knob to review the	
	function		ime and use the knob to review the most recent historical waveforms can	
Regulatory compliance		data. When freezing, 30 i		
Regulatory compliance	function	data. When freezing, 30 i	most recent historical waveforms can	
· , .	function should conform to the	data. When freezing, 30 is be reviewed by sliding th	most recent historical waveforms can e screen or rotating the knob.	
	function should conform to the requirements of YY0784-2010	data. When freezing, 30 is be reviewed by sliding th	most recent historical waveforms can e screen or rotating the knob. Up to 5 reference loops can be	
· , .	function should conform to the requirements of YY0784-2010 Nellcor SpO2: (Lower alarm limit	data. When freezing, 30 is be reviewed by sliding the Freeze the loop review	most recent historical waveforms can e screen or rotating the knob. Up to 5 reference loops can be	
· , .	function should conform to the requirements of YY0784-2010 Nellcor SpO2: (Lower alarm limit +1%)~100%	data. When freezing, 30 is be reviewed by sliding the Freeze the loop review O ₂ Therapy	most recent historical waveforms can e screen or rotating the knob. Up to 5 reference loops can be saved.	
	function should conform to the requirements of YY0784-2010 Nellcor SpO2: (Lower alarm limit +1%)~100% Masimo SpO2: (Lower alarm limit	data. When freezing, 30 mbe reviewed by sliding the Freeze the loop review O2 Therapy O2 %	most recent historical waveforms can e screen or rotating the knob. Up to 5 reference loops can be saved. 21 - 100 % (increments of 1 %) ± (3	
· , .	function should conform to the requirements of YY0784-2010 Nellcor SpO2: (Lower alarm limit +1%)~100% Masimo SpO2: (Lower alarm limit +1%)~100%	data. When freezing, 30 mbe reviewed by sliding the Freeze the loop review O2 Therapy O2 %	most recent historical waveforms can e screen or rotating the knob. Up to 5 reference loops can be saved. 21 - 100 % (increments of 1 %) ± (3 vol.% +1 % of setting)	
Jpper SpO2 alarm limit	function should conform to the requirements of YY0784-2010 Nellcor SpO2: (Lower alarm limit +1%)~100% Masimo SpO2: (Lower alarm limit +1%)~100% Comen SpO2: (Lower alarm limit	data. When freezing, 30 mbe reviewed by sliding the Freeze the loop review O2 Therapy O2 % Flow2 - 60 L/min ± (1.5 L) Gas Circuit	most recent historical waveforms can e screen or rotating the knob. Up to 5 reference loops can be saved. 21 - 100 % (increments of 1 %) ± (3 vol.% +1 % of setting)	
Jpper SpO2 alarm limit	function should conform to the requirements of YY0784-2010 Nellcor SpO2: (Lower alarm limit +1%)~100% Masimo SpO2: (Lower alarm limit +1%)~100% Comen SpO2: (Lower alarm limit +1%)~100%	data. When freezing, 30 mbe reviewed by sliding the Freeze the loop review O ₂ Therapy O ₂ % Flow2 - 60 L/min ± (1.5 L) Gas Circuit Specification	most recent historical waveforms can e screen or rotating the knob. Up to 5 reference loops can be saved. 21 - 100 % (increments of 1 %) ± (3 vol.% +1 % of setting) /min +10 % of setting) (BTPS)	
Jpper SpO2 alarm limit	function should conform to the requirements of YY0784-2010 Nellcor SpO2: (Lower alarm limit +1%)~100% Masimo SpO2: (Lower alarm limit +1%)~100% Comen SpO2: (Lower alarm limit +1%)~100% Nellcor SpO2: 20%~(Upper alarm	data. When freezing, 30 mbe reviewed by sliding the Freeze the loop review O2 Therapy O2 % Flow2 - 60 L/min ± (1.5 L) Gas Circuit Specification Gas type	most recent historical waveforms can e screen or rotating the knob. Up to 5 reference loops can be saved. 21 - 100 % (increments of 1 %) ± (3 vol.% +1 % of setting) /min +10 % of setting) (BTPS)	
Upper SpO2 alarm limit	function should conform to the requirements of YY0784-2010 Nellcor SpO2: (Lower alarm limit +1%)~100% Masimo SpO2: (Lower alarm limit +1%)~100% Comen SpO2: (Lower alarm limit +1%)~100% Nellcor SpO2: 20%~(Upper alarm limit -1%)	data. When freezing, 30 mbe reviewed by sliding the Freeze the loop review O2 Therapy O2 % Flow2 - 60 L/min ± (1.5 L) Gas Circuit Specification Gas type Gas source requirement	most recent historical waveforms can e screen or rotating the knob. Up to 5 reference loops can be saved. 21 - 100 % (increments of 1 %) ± (3 vol.% +1 % of setting) /min +10 % of setting) (BTPS)	
Upper SpO2 alarm limit	function should conform to the requirements of YY0784-2010 Nellcor SpO2: (Lower alarm limit +1%)~100% Masimo SpO2: (Lower alarm limit +1%)~100% Comen SpO2: (Lower alarm limit +1%)~100% Nellcor SpO2: 20%~(Upper alarm limit -1%) Masimo SpO2: 1%~(Upper alarm	data. When freezing, 30 mbe reviewed by sliding the Freeze the loop review O2 Therapy O2 % Flow2 - 60 L/min ± (1.5 L) Gas Circuit Specification Gas type	most recent historical waveforms can e screen or rotating the knob. Up to 5 reference loops can be saved. 21 - 100 % (increments of 1 %) ± (3 vol.% +1 % of setting) /min +10 % of setting) (BTPS)	
Regulatory compliance Upper SpO2 alarm limit Lower SpO2 alarm limit	function should conform to the requirements of YY0784-2010 Nellcor SpO2: (Lower alarm limit +1%)~100% Masimo SpO2: (Lower alarm limit +1%)~100% Comen SpO2: (Lower alarm limit +1%)~100% Nellcor SpO2: 20%~(Upper alarm limit -1%) Masimo SpO2: 1%~(Upper alarm limit -1%)	data. When freezing, 30 mbe reviewed by sliding the Freeze the loop review O2 Therapy O2 % Flow2 - 60 L/min ± (1.5 L) Gas Circuit Specification Gas type Gas source requirement	most recent historical waveforms can e screen or rotating the knob. Up to 5 reference loops can be saved. 21 - 100 % (increments of 1 %) ± (3 vol.% +1 % of setting) /min +10 % of setting) (BTPS)	



range			collection cup, expiratory valve,
Rated flow rate requireme			Y-joint): ≤ 2 mL/cmH2O;
Input connector	NIST (ISO 5356-1) or DISS (CGA 1240)		Pediatric disposable circuit
Standards compliant	YY/T 0799-2010		(including inspiratory safety valve,
	EN ISO5359:2008		pediatric disposable breathing
Low-pressure O ₂ source			tube, water collection cup,
Input pressure range	< 100 kPa		expiratory valve): ≤ 2 mL/cmH2O;
Maximum flow rate 15 l	_/min		Pediatric reusable circuit (including
Input connector	CPC quick coupling		inspiratory safety valve, pediatric
Inspiratory module			reusable breathing tube, water
Peak flow rate	≥ 210 L/min		collection cup, expiratory valve,
Nebulizer connectorFlow	rate: 5 L/min \sim 8 L/min		Y-joint): ≤ 2 mL/cmH2O;
Safety pressure of	≤ 12.5 kPa		Neonate reusable circuit (including
respiration			inspiratory safety valve, neonate
Inspiratory-side external connector	Coaxial 22 mm/15 mm conical connector		reusable breathing tube, water
Removable, sterilizable	can be entirety removed quickly;		collection cup, expiratory valve,
,	and can be entirety cleaned and		Y-joint): ≤ 1 mL/cmH2O.
	disinfected.	Inspiratory resistance≤ 6 cmH2O at the flow rate of 60 L/mir (Adult);	
Regulatory compliance	YY1040.1-2003		
	EN ISO5356-1:2004	≤ 6 cmH2O at the flow rate	te of 30 I /min (Pediatric):
Expiratory module		≤ 6 cmH2O at the flow rate	
Expiratory-side external	Coaxial 22 mm/15 mm conical	Expiratory resistance	\leq 6 cmH2O at the flow rate of 60
connector	connector	Expiratory resistance	
Removable, sterilizableca	n be entirety removed quickly; and		L/min (Adult);
can be entirety cleaned ar	nd disinfected.		≤ 6 cmH2O at the flow rate of 30
Regulatory compliance	YY1040.1-2003		L/min (Pediatric);
	EN ISO5356-1:2004		≤ 6 cmH2O at the flow rate of 5
System compliance and			L/min (Neonate).
resistance		Basic performance	
Compliance	Adult disposable circuit (including	Pressure monitoring range	-20~120 cmH2O
	inspiratory safety valve, adult		
	disposable breathing tube, water	valve): ≤ In non-ventilation state or power failure or gas source failure (<0.12MPa): < 14 cmH2O	
	collection cup, expiratory valve): ≤		
	4 mL/cmH2O;		
	Adult reusable circuit (including		
	inspiratory safety valve, adult	Environmental	



Temperature	5 - 40 °C (operating); -20 to 60 °C (storage and transport, O₂ sensor: -20 to 50°C)	Battery life	140 min (when a new fully charged battery is used in standard operating mode)
Relative Humidity5 - 95 % transport)	(operating); 5 - 95 % (storage and		280 min (when two new fully charged batteries are used in
Barometric Pressure	62 - 106 kPa (operating); 50 -106 kPa (storage and transport)		standard operating mode)
Power Specification			
External AC power supply			
Input voltage100 - 240 V			
Input frequency50/60 Hz			
Input current	1.2 – 0.5 A		
Fuse	T3AL/250 V		
External DC power			
supply		Recharge Time	Single battery:3 hours
Input voltage	12V	1/0	
Input current10A		Communication	Rs232, Ethernet, VGA, USB port,
Internal battery		interface	Nurse call
Number of batteries	One or Two	Trolley MC100	
Battery typeBuild-in Lithiu	um-ion battery, 14.4 VDC, 6700mAh	Dimensions	687.5×528.8×1040.6 mm
		Weight	20 kg

Notice: Specifications subject to changes without prior

notice. All rights reserved by Comen